



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ANALYTICAL & PRECISION BALANCE CO., INC.
9830 South 51st Street, Suite B-103
Phoenix, AZ 85044
Nicole Brady Phone: 1 480 598 0321

CALIBRATION

Valid To: July 31, 2028

Certificate Number: 7942.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical & Mechanical Balances ³	(1.2 to 120) g (120 to 210) g (210 to 600) g (600 to 1200) g 1200 g to 5 kg (5 to 25) kg (25 to 50) kg (50 to 70) kg (70 to 120) kg (120 to 160) kg (160 to 320) kg	0.049 mg 0.29 mg 0.59 mg 0.60 mg 5.9 mg 0.24 g 0.34 g 0.42 g 0.58 g 0.67 g 0.94 g	Calibration procedure QP-035 (direct method) by using Class 1 weights, Class F weights
Platform Scales ³	(25 to 5000) lb (5000 to 10 000) lb	0.068 lb 0.076 lb	Calibration procedure QP-035, (direct method) by using Class F weights

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the Calibration and Measurement Capability Uncertainty (CMC) found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



Accredited Laboratory

A2LA has accredited

ANALYTICAL & PRECISION BALANCE CO., INC.

Phoenix, AZ

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 3rd day of August 2026.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 7942.01
Valid to July 31, 2028
Revised August 7, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.